


VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


5/3/10
NJB 6/9/10
5/4/10
4/29/10 CCB
5/3/10
CCB 5/24/10


Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130 286
Name	Stan Swetzof	ID#	5002
Date	5/3/10		
Agency	North Pole Police Department	Phone #	488-0612
Instrument Location	125 Shawman Lane, North Pole, AK 99705		
Alco S/N	X172 959	Target Value	.083
		High Pressure	375 PSI
Alco Test Values	.079	.079	
	1 st Alco	2 nd Alco	
Signature			
(OVER)			


(Do Not write in the area below)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 10th day of June, 2010.

(Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130286

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130286

MAY 03, 2010

OPERATOR'S NAME:

SMETZOF-STANLEY

OPERATOR'S NUMBER: 5002

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

L

O.L. #: L

DEPT/AGENCY: NPL1

CASE/REPORT:

TEST TYPE: U

ALCO TARGET VALUE: .003

ALCO S/N: X172959

--- BREATH ANALYSIS ---

.003 ADJUSTED FOR 29.96 in
ALCO TARGET .003 12:30
BLANK TEST .000 12:31
INTERNAL STANDARD VERIFIED 12:31
ALCO TV 29.96 in .079 12:31
BLANK TEST .000 12:32
SUBJECT SAMPLE .000 12:33
BLANK TEST .000 12:34
ALCO TV 29.98 in .079 12:34
BLANK TEST .000 12:35

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130286

MAY 03, 2010

TIME 12:25

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/00/09
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 42c
BAROMETER: 29.98 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=>?@ABCDEFGHIJ
KLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~■